




America

CERTIFICATE

No. U8 111568 0001 Rev. 00

Holder of Certificate: **nevoLAB GmbH**
Am Gehrenbach 8
88167 Maierhöfen
GERMANY

Certification Mark:



Product: **Laboratory Equipment**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713207470

Date, 2021-08-27

(Ralph Fischer)



America

CERTIFICATE

No. U8 111568 0001 Rev. 00

Model(s): XELSIUS Reactor

Brand Name: nevoLAB

Tested according to: UL 61010-1:2012/R:2019-07
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11
UL 61010-2-010:2019
CAN/CSA-C22.2 No. 61010-2-010:2019
UL 61010-2-051:2019
CAN/CSA-C22.2 No. 61010-2-051:2019

Parameters:

Rated voltage:	240 VAC
Rated frequency:	50/60 Hz
Rated power:	max. 1.4 kW
Protection class:	I